

Precision Chiropractic

Steven Danaher, D.C.
712 D Street, Suite J, San Rafael, CA 94901
www.MarinNucca.com

Imprinter Cleaning and Testing

Cleaning

- Use a cotton ball plus some alcohol.
- Wipe down the top surface, all four sides, and the bottom.
- Use a moderate amount of alcohol, do not pour alcohol over the unit as alcohol will drag contaminants down into the inside of the imprinter, rendering it ineffective.
- Use cotton, not a paper towel, as paper will pull the lettering off of the imprinter.
- Do this at least once a month, sometimes once per week.
- Disinfecting with a cold laser provides very short term results, best to use alcohol.



Battery Testing

- The imprinter works best when the batteries measure at least 8.5 volts.
- Best to remove the batteries and test with a digital voltmeter.
- When new the batteries measure 9.3 volts or greater. At 8.5 volts the battery is 95% depleted.

